



Certificate of Compliance

Certificate: 1088063 (LR65844C)

Master Contract: 176045

Project: 1529928

Date Issued: 2004/03/12

Issued to: Carl Zeiss Jena GmbH
Pruefzentrum
Carl-Zeiss-Promenade 10
Jena, 07745
Germany
Attention: Dr. Wolfram Daenhardt

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by: Peter Cheung, P. Eng.

Authorized by: Ray Fadavi, P.Eng., MBA,
Operations Manager

PRODUCTS

CLASS 8750 81 - MEDICAL ELECTRICAL EQUIPMENT - Certified to US Standards
CLASS 8750 01 - MEDICAL ELECTRICAL EQUIPMENT

Laser scanning microscope, Model LSM510, and LSM510 META, cord-connected, rated 120V and 240 V, 50/60Hz, 5400VA, Pollution Degree 2, Installation Category II.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 1010.1-92 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 1010.1B-92 - Amendment 2: 1997 to CAN/CSA-C22.2 No. 101.1-92, Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



CSA INTERNATIONAL

Certificate: 1088063 (LR65844C)

Master Contract: 176045

Project: 1529928

Date Issued: 2004/03/12

UL Std No. 61010A-1

- Electrical Equipment for Laboratory Use; Part 1: General Requirements